

Certificate of Traceability

Level I Laboratory Certification

certification and is traceable to the Diabetes Control and Complications Trial Reference method. The comparison was performed with: University of Missouri SRL#12 Variant II has participated in and successfully completed NGSP Level I Laboratory This certifies that Eurofins Central Laboratory Pte Ltd, using Bio-Rad

The system evaluated was:

Instrument:	Calibrator Lot:	Column Lot:
Variant II	64594144, 64594147	00061
Reagent Lot:	Calibrator Assigned Values:	
64596771, 64596772,64594609	5.6%, 10.7%	

Date of Certification: July 1, 2025

Certification Expires: July 1, 2026

Cut Rolling

NGSP Steering Committee Chair

Sand South

NGSP Network Coordinator

SRL director/ supervisor

Charda Hancod